

Results 6th Gravimetric Interlaboratory Test 2025



Part.-No.	Tube-No.	empty	filled	difference [g]	difference [mg]	determined [mg]	Result [mg]	Remark	Recovery rate [%]
2025-0001	11	0,81061	0,81514	0,00453	4,53000	4,77000	-0,24000		105,3
	12	0,78109	0,78606	0,00497	4,97000	5,31000	-0,34000		106,8
	13	0,77447	0,77987	0,00540	5,40000	6,57000	-1,17000		121,7
	14	0,81848	0,82371	0,00523	5,23000	5,64000	-0,41000		107,8
2025-0002	15	0,81040	0,81599	0,00559	5,59000		5,59000	Result not submitted	0,0
	16	0,80673	0,81202	0,00529	5,29000		5,29000	Result not submitted	0,0
	17	0,79468	0,799761	0,00508	5,08100		5,08100	Result not submitted	0,0
	18	0,76487	0,771178	0,00631	6,30800		6,30800	Result not submitted	0,0
2025-0003	19	0,77204	0,77755	0,00551	5,51000		5,51000	Result not submitted	0,0
	20	0,79140	0,79698	0,00558	5,58000		5,58000	Result not submitted	0,0
	21	0,80586	0,81071	0,00485	4,85000		4,85000	Result not submitted	0,0
	22	0,80545	0,81045	0,00500	5,00000		5,00000	Result not submitted	0,0
2025-0004	23	0,77524	0,7808	0,00556	5,56000		5,56000	Result not submitted	0,0
	24	0,79630	0,80108	0,00478	4,78000		4,78000	Result not submitted	0,0
	25	0,79951	0,80357	0,00406	4,06000		4,06000	Result not submitted	0,0
	26	0,78294	0,7893	0,00636	6,36000		6,36000	Result not submitted	0,0
2025-0005	27	0,79702	0,80325	0,00623	6,23000		6,23000	Result not submitted	0,0
	28	0,76516	0,77009	0,00493	4,93000		4,93000	Result not submitted	0,0
	29	0,80624	0,81238	0,00614	6,14000		6,14000	Result not submitted	0,0
	30	0,79830	0,80396	0,00566	5,66000		5,66000	Result not submitted	0,0
2025-0006	31	0,79514	0,80161	0,00647	6,47000		6,47000	Result not submitted	0,0
	32	0,79581	0,80117	0,00536	5,36000		5,36000	Result not submitted	0,0
	33	0,73989	0,74439	0,00450	4,50000		4,50000	Result not submitted	0,0
	34	0,79924	0,80452	0,00528	5,28000		5,28000	Result not submitted	0,0
2025-0007	35	0,81109	0,81597	0,00488	4,87900		4,87900	Result not submitted	0,0
	36	0,79117	0,79653	0,00536	5,36000		5,36000	Result not submitted	0,0
	37	0,76633	0,77276	0,00643	6,43000		6,43000	Result not submitted	0,0
	38	0,79166	0,79705	0,00539	5,39000		5,39000	Result not submitted	0,0
2025-0008	39	0,79823	0,8034	0,00517	5,17000		5,17000	Result not submitted	0,0
	40	0,78910	0,79372	0,00462	4,62000		4,62000	Result not submitted	0,0
	41	0,78440	0,78983	0,00543	5,43000		5,43000	Result not submitted	0,0
	42	0,81544	0,82029	0,00485	4,85000		4,85000	Result not submitted	0,0
2025-0009	43	0,76313	0,76867	0,00554	5,54000	5,50000	0,04000		99,3
	44	0,80041	0,80711	0,00670	6,70000	6,76000	-0,06000		100,9
	45	0,77428	0,77924	0,00496	4,96000	5,07000	-0,11000		102,2
	46	0,82929	0,83514	0,00585	5,85000	5,87000	-0,02000		100,3
2025-0010	47	0,76651	0,77331	0,00680	6,80000	6,60000	0,20000		97,1
	48	0,81330	0,81884	0,00554	5,54000	5,80000	-0,26000		104,7
	49	0,78286	0,78977	0,00691	6,91000	5,50000	1,41000		79,6
	50	0,76482	0,76953	0,00471	4,71000	3,90000	0,81000		82,8
2025-0011	51	0,79779	0,80215	0,00436	4,36000	4,41000	-0,05000		101,1
	52	0,76510	0,7718	0,00670	6,70000	6,76000	-0,06000		100,9
	53	0,79611	0,80224	0,00613	6,13000	6,12000	0,01000		99,8
	54	0,78942	0,79483	0,00541	5,41000	5,44000	-0,03000		100,6
2025-0012	55	0,78090	0,78695	0,00605	6,05000		6,05000	Result not submitted	0,0
	56	0,78430	0,78986	0,00556	5,56000	4,37000	1,19000		78,6
	57	0,78313	0,7894	0,00627	6,27000		6,27000	Result not submitted	0,0
	58	0,80358	0,80822	0,00464	4,64000	5,45000	-0,81000		117,5
2025-0013	59	0,79530	0,80038	0,00508	5,08000		5,08000	Result not submitted	0,0
	60	0,77246	0,77911	0,00665	6,65000		6,65000	Result not submitted	0,0
	61	0,83917	0,84478	0,00561	5,61000		5,61000	Result not submitted	0,0
	62	0,82365	0,82919	0,00554	5,54000		5,54000	Result not submitted	0,0
2025-0014	63	0,76512	0,7725	0,00738	7,38000		7,38000	Result not submitted	0,0
	64	0,77629	0,78358	0,00729	7,29000		7,29000	Result not submitted	0,0
	65	0,78726	0,79404	0,00678	6,78000		6,78000	Result not submitted	0,0
	66	0,77299	0,77815	0,00516	5,16000		5,16000	Result not submitted	0,0
2025-0015	67	0,79571	0,80259	0,00688	6,88000	6,87000	0,01000		99,9
	68	0,79147	0,79614	0,00467	4,67000	4,74000	-0,07000		101,5
	69	0,79542	0,80107	0,00565	5,65000	5,70000	-0,05000		100,9
	70	0,82285	0,82833	0,00548	5,48000	5,74000	-0,26000		104,7
2025-0016	71	0,79983	0,80502	0,00519	5,19000	5,23000	-0,04000		100,8
	72	0,81077	0,8166	0,00583	5,83000	6,06000	-0,23000		103,9
	73	0,78894	0,79474	0,00580	5,80000	5,91000	-0,11000		101,9
	74	0,78336	0,78937	0,00601	6,01000	6,17000	-0,16000		102,7
2025-0017	75	0,80247	0,80977	0,00730	7,30000	7,00000	0,30000		95,9
	76	0,81826	0,82343	0,00517	5,17000	5,24000	-0,07000		101,4
	77	0,78427	0,79094	0,00667	6,67000	6,65000	0,02000		99,7
	78	0,77252	0,77937	0,00685	6,85000	6,75000	0,10000		98,5
2025-0018	79	0,79576	0,80151	0,00575	5,75000	4,60000	1,15000		80,0
	80	0,79557	0,80147	0,00590	5,90000	5,20000	0,70000		88,1
	81	0,74004	0,74663	0,00659	6,59000	6,30000	0,29000		95,6
	82	0,76197	0,76702	0,00505	5,05000	4,40000	0,65000		87,1

Results 6th Gravimetric Interlaboratory Test 2025



Part.-No.	Tube-No.	empty	filled	difference [g]	difference [mg]	determined [mg]	Result [mg]	Remark	Recovery rate [%]
2025-0019	83	0,80012	0,80592	0,00580	5,80000	5,62000	0,18000		96,9
	84	0,80195	0,80675	0,00480	4,80000	4,54000	0,26000		94,6
	85	0,80664	0,81247	0,00583	5,83000	5,79000	0,04000		99,3
	86	0,76684	0,77357	0,00673	6,73000	6,70000	0,03000		99,6
2025-0020	87	0,81344	0,822	0,00856	8,56000	8,49000	0,07000		99,2
	88	0,79223	0,79715	0,00492	4,92000	5,08000	-0,16000		103,3
	89	0,79141	0,7959	0,00449	4,49000	4,55000	-0,06000		101,3
	90	0,81771	0,8234	0,00569	5,69000	5,81000	-0,12000		102,1
2025-0021	91	0,79260	0,80054	0,00794	7,94000	7,67000	0,27000		96,6
	92	0,78353	0,78788	0,00435	4,35000	4,80000	-0,45000		110,3
	93	0,78328	0,78849	0,00521	5,21000	5,34000	-0,13000		102,5
	94	0,78680	0,79034	0,00354	3,54000	3,73000	-0,19000		105,4
2025-0022	95	0,78254	0,78914	0,00660	6,60000	5,92000	0,68000		89,7
	96	0,81551	0,82198	0,00647	6,47000	6,48000	-0,01000		100,2
	97	0,82354	0,82876	0,00522	5,22000	5,32000	-0,10000		101,9
	98	0,81082	0,81639	0,00557	5,57000	5,21000	0,36000		93,5
2025-0023	99	0,80936	0,81396	0,00460	4,60000	4,80000	-0,20000		104,3
	100	0,79461	0,80093	0,00632	6,32000	4,70000	1,62000		74,4
	101	0,78566	0,79076	0,00510	5,10000	2,90000	2,20000		56,9
	102	0,79722	0,8028	0,00558	5,58000	4,20000	1,38000		75,3
2025-0024	103	0,79118	0,79593	0,00475	4,75000	2,20000	2,55000		46,3
	104	0,77408	0,78008	0,00600	6,00000	4,50000	1,50000		75,0
	105	0,76670	0,7731	0,00640	6,40000	4,10000	2,30000		64,1
	106	0,78282	0,7895	0,00668	6,68000	4,20000	2,48000		62,9
2025-0025	107	0,79251	0,79853	0,00602	6,02000	5,14000	0,88000		85,4
	108	0,79902	0,80418	0,00516	5,16000	4,83000	0,33000		93,6
	109	0,78389	0,79015	0,00626	6,26000	5,89000	0,37000		94,1
	110	0,79920	0,80414	0,00494	4,94000	4,98000	-0,04000		100,8
2025-0026	111	0,80422	0,80966	0,00544	5,44000	4,80000	0,64000		88,2
	112	0,76629	0,77271	0,00642	6,42000	5,70000	0,72000		88,8
	113	0,82951	0,83432	0,00481	4,81000	3,90000	0,91000		81,1
	114	0,80692	0,81221	0,00529	5,29000	4,80000	0,49000		90,7
2025-0027	115	0,80247	0,8097	0,00723	7,23000	5,51000	1,72000		76,2
	116	0,79046	0,79536	0,00490	4,90000	5,89000	-0,99000		120,2
	117	0,76244	0,76866	0,00622	6,22000	5,75000	0,47000		92,4
	118	0,79598	0,80121	0,00523	5,23000	2,94000	2,29000		56,2
2025-0028	119	0,76653	0,7728	0,00627	6,27000	6,62000	-0,35000		105,6
	120	0,80992	0,81617	0,00625	6,25000	5,56000	0,69000		89,0
	121	0,80977	0,8156	0,00583	5,83000	5,33000	0,50000		91,4
	122	0,77427	0,77928	0,00501	5,01000	4,26000	0,75000		85,0
2025-0029	123	0,80239	0,8081	0,00571	5,71000	5,71000	0,00000		100,0
	124	0,78936	0,79413	0,00477	4,77000	4,77000	0,00000		100,0
	125	0,78938	0,7954	0,00602	6,02000	6,02000	0,00000		100,0
	126	0,78282	0,78888	0,00606	6,06000	6,05000	0,01000		99,8
2025-0030	127	0,82198	0,82827	0,00629	6,29000	2,12000	4,17000		33,7
	128	0,78800	0,79434	0,00634	6,34000	1,54000	4,80000		24,3
	129	0,81048	0,8157	0,00522	5,22000	1,10000	4,12000		21,1
	130	0,78726	0,79195	0,00469	4,69000	1,01000	3,68000		21,5